

<b>X24A SAFETY CHECKLIST</b>			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
<b>Rev: E</b>	<b>Effective: 5/24/2010</b>	<b>Page 1 of 1</b>	Number: LS-SCL-0054
Prepared by: Philip Marino		Reviewed by: P. Marino	<b>Approved by: B. Chmiel</b>

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

- 1. Bremsstrahlung shield (BS #1) between GV1 and Mirror Box #1 is banded as per photo.
- 2. Exclusion Zone # 1 (EZ#1) between monochromator and GV2, defined by polystyrene is identified and secured.
- 3. The 6" viewport located on the side of monochromator tank is covered by lead glass.
- 4. Exclusion Zone # 2 (EZ#2) between monochromator and GV4, defined by polystyrene is identified and secured.
- 5. Bremsstrahlung shield (BS # 2) between the monochromator and GV4 in place and banded as per photo.

OPCO / Beamline Staff: \_\_\_\_\_ Date: \_\_\_\_\_