

Beamline Hazard Analysis - Beamline X23A2

LS-BHA-X23A2

Revision: 01

Effective:

Page 1 of 2

The following tasks are included in the beamline safety review. Any operations not listed here, or performed by others than those listed, need additional review by ES&H personnel and/or beamline personnel.

Prepared By: B. Ravel

Approved By: Lori Stiegler

Tasks / Equipment	Performed by:		Controls
	Beamline Staff	Users	
Inert compressed gas	Y	Y	<ul style="list-style-type: none"> Compressed gas training – TQ-COMPGAS1 or NSLS User training Transport cylinders using a cylinder cart Properly secure cylinders Wear safety glasses with side shields while connecting lines Leak check gas lines
Toxic compressed gases	Y	Y	<ul style="list-style-type: none"> Compressed gas training – TQ-COMPGAS1 or NSLS User training Transport cylinders using a cylinder cart Properly secure cylinders Use small quantities or a ventilated gas cabinet Wear safety glasses with side shields while connecting lines Leak check gas lines Small quantities limited by SAF Ventilation system in hutch
Flammable compressed gases	Y	Y	<ul style="list-style-type: none"> Compressed gas training – TQ-COMPGAS1 or NSLS User training Transport cylinders using a cylinder cart Properly secure cylinders Use small quantities or a ventilated gas cabinet Use a flash arrestor Wear safety glasses with side shields while connecting lines Leak check gas lines
Solvents	Y	Y	<ul style="list-style-type: none"> Limit quantities at beamline to < 250 ml Use PPE - safety glasses, nitrile gloves
Pouring cryogen ≤ 5L	Y	Y	<ul style="list-style-type: none"> Use goggles Long pants and sleeves - avoid clothing which can trap cryogen (pockets, cuffs) Cryogen or heavy gloves Enclosed shoes Never pour from above chest level
Using the LN2 Fill Station or pressurized transfer	Y	Y	<ul style="list-style-type: none"> LS-PROC-LN2FILL-0709 training for Auto station Use full face shield along with goggles or safety

Beamline Hazard Analysis - Beamline X23A2

LS-BHA-X23A2

Revision: 01

Effective:

Page 2 of 2

The following tasks are included in the beamline safety review. Any operations not listed here, or performed by others than those listed, need additional review by ES&H personnel and/or beamline personnel.

			<ul style="list-style-type: none"> glasses • Long pants and sleeves - avoid clothing which can trap cryogenics (pockets, cuffs) • Cryogen or heavy gloves • Enclosed shoes
Manipulating cryogenic samples	Y	Y	<ul style="list-style-type: none"> • Use goggles • Long pants • Impervious disposable gloves over cotton liners • Enclosed shoes • Use tongs or tools whenever possible
Soldering	Y	N	<ul style="list-style-type: none"> • Use a designated area, periodically clean surfaces
Working at heights	Y	N	<ul style="list-style-type: none"> • Be aware of body posture, training
Beryllium use	Y	Y	<ul style="list-style-type: none"> • Handle articles with gloves, • Leave area and contact OpCo if damaged